

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE  
CERTIFICAT D'ESSAI OC

Product  
Produit

AC Power Adapter

Name and address of the applicant  
Nom et adresse du demandeur

ICCNexergy Inc - Design Center  
5th Floor western,  
Changhua Building, No. 921 Xingye Road, Nancun Town, Panyu,  
Guangzhou, 511442 Guangdong, China

Name and address of the manufacturer  
Nom et adresse du fabricant

ICCNexergy, Inc.  
4 Westbrook Corporate Center, SUITE 900  
Westchester IL 60154, USA

Name and address of the factory  
Nom et adresse de l'usine

ICC Electronics (Dongguan) Ltd.  
No. 23, Shang Yuan Road  
Qing Xi Town, Dong Guan City, 523640 Guangdong, China

Note: When more than one factory, please report on page 2  
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2<sup>ème</sup> page

Ratings and principal characteristics  
Valeurs nominales et caractéristiques principales

Input: AC 100-240V; 50-60Hz; 1.5-0.6A;  
Output: DC 19V; 5.26A; 100W max

Trade mark (if any)  
Marque de fabrique (si elle existe)

INVACARE

Model/type Ref.  
Ref. de type

MWC100019B  
POC1-13X-YY (X=0-9,Y=Blank or A-Z)

Additional information (if necessary may also be  
reported on page 2)  
Les Information complémentaire (si nécessaire,  
peuvent être indiqués sur la 2<sup>ème</sup> page)

-- The risk management requirements of the standard were not  
addressed.

A sample of the product was tested and found  
to be in conformity with  
Un échantillon de ce produit a été essayé et a été  
considéré conforme à la

PUBLICATION

EDITION

IEC 60601-1:2005 + Corrigendum 1 + Corrigendum 2 + A1

As shown in the Test Report Ref. No. which forms part  
of this Certificate  
Comme indiqué dans le Rapport d'essais numéro de  
référence qui constitue une partie de ce Certificat

17047121 001

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme National de Certification



TÜV Rheinland LGA Products GmbH  
Tillystraße 2 · 90431 Nürnberg, Germany  
Phone + 49 221 806-1371  
Fax + 49 221 806-3935  
Mail: cert-validity@de.tuv.com  
Web: www.tuv.com



*C. Nasca*

Date: 05.05.2015

Signature:

Dipl.-Ing. (FH) C. Nasca

ICCNexergy Inc-Design Center

Date : 2015-05-05

Mr. Andy Quan

Our ref. : awa ZD

Your ref. : 164032129

5th Floor western, Changhua Building,  
No.921 Xingye Road, Nancun Town, Panyu,  
Guangzhou, Guangdong 511442, P.R.China

**Ref : CB Certificate Germany**

Type of Equipment : AC Power Adapter

Model Designation: See Certificate

Certificate No. : DE 2-018858

Report No. : 17047121 001

Dear Mr. Andy Quan,


Thank you very much for your interest in our services.

Please find enclosed your certification documents.

We appreciate your support and would like to offer our assistance in the approval of your future products through our extensive range of technical services. Please feel free to contact us whatever your requirements may be.

With kind regards,

Certification Body



Dipl.-Ing. (FH) C. Nasca

Enclosure

证书的详细资料请登陆[www.tuvdotcom.com](http://www.tuvdotcom.com)查阅,或拨打我司客服热线800 999 3668 / 400 883 1300咨询