



Ref. Certif. No.

DE 2-024275

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Power Supply

Name and address of the applicant

Inventus Power, Inc. - Technical Center
5th Floor Western, ChangHua Building
No. 921 XingYe Road, NanCun Town, Panyu, Guangzhou City,
511442 Guangdong, China

Name and address of the manufacturer

Inventus Power, Inc.
1200 Internationale Parkway
Woodridge IL 60517, USA

Name and address of the factory

ICC Electronics (Dongguan) Ltd.
No. 23, Shang Yuan Road
Qing Xi Town, Dongguan City, 523640 Guangdong, China

Ratings and principal characteristics

Input: 100-240Vac, 50-60Hz, 1.0-0.6A
Output: Refer to general product information for details.

Trademark (if any)

ELPAC POWER SYSTEMS

Customer's Testing Facility (CTF) Stage used

N/A

Model / Type Ref.

MTA040xxxz
xxx=009 to 024
z=A to Z or blank, not safety related

Additional information (if necessary may also be reported on page 2)

The risk management requirements of the standard were not addressed.

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005+A1
for national deviations see test report

As shown in the Test Report Ref. No. which forms part of this Certificate

50202164 001

This CB Test Certificate is issued by the National Certification Body



TÜV Rheinland LGA Products GmbH
Tillystraße 2 · 90431 Nürnberg, Germany
Phone + 49 221 806-1371
Fax + 49 221 806-3935
Mail: cert-validity@de.tuv.com
Web: www.tuv.com



Date: 11.12.2018

Signature:

Dipl.-Ing. Univ. S. O. Steinke