

# Certificate



Certificate no.

CU 72130841 01

**License Holder:**

ICCNexergy, Inc.  
4 Westbrook Corp. Center, Ste. 900  
Westchester IL 60154  
USA

**Manufacturing Plant:**

ICC Electronics (Dongguan) Ltd.  
No. 23, Shang Yuan Road, Qingxi Town,  
Dong Guan City, Guangdong  
China

Test report no.: USA-RF 30682009 007

Client Reference: Carol Pickford

Tested to: ANSI/AAMI ES60601-1:2005  
CAN/CSA-C22.2 No. 60601-1-08

**Certified Product:** Medical Switch Mode Power Supply

**License Fee - Units**

Model Designation:

- 1) MSMP250xxz, ELPAC POWER SYSTEMS MSMP250xxz;
- 2) FXA3500xxA, ELPAC POWER SYSTEMS FXA3500xxA  
(xx = 12-48; z = A-Z, blank; not safety-relevant)

7

Rated Voltage: AC 100-240V, 50/60Hz

Rated Current: 5A

Protection Class: I


Output Ratings: DC 12-48V

- 1) 20A max., 250W max. (with convection cooling),  
28A max., 350W max. (with 26CFM min. airflow);
- 2) 28A max., 350W max. (with 26CFM min. airflow)

Appendix: 1, 1-11

7

Licensed Test mark:



Date of Issue

(day/mo/yr)

12/04/2013